

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 26 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130184</u>	
Name <u>SGT. HAMILTON</u>		ID# <u>3192</u>	Date <u>4/21/10</u>
A	Agency <u>PETERSBURG POLICE</u>		Phone # <u>772-3838</u>
Instrument Location <u>PETERSBURG</u>			
Alco S/N <u>78268 Smb</u> <u>X78268</u>		Target Value <u>.081</u>	High Pressure <u>200 PSI</u>
B	Alco Test Values		
	<u>.080</u> 1 <sup>st</sup> Alco	<u>.081</u> 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>			
(OVER)			

Smb  
4/28/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 30th day of April, 2010.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

APRIL 21, 2010

OPERATOR'S NAME:  
HAMILTON JOHN A  
OPERATOR'S NUMBER: 3192  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI:

A  
O.L. #: A  
DEPT/AGENCY: PSP01  
CASE/REPORT: A  
TEST TYPE: U  
ALCO TARGET VALUE: 0.08  
0.00-0.07 476268

--- BREATH ANALYSIS ---  
0.081 ADJUSTED FOR 30.03 in 06:11  
ALCO TARGET .081 06:13  
BLANK TEST .000 06:13  
INTERNAL STANDARD VERIFIED 06:13  
ALCO TO 30.03 in .000 06:13  
BLANK TEST .000 06:14  
SUBJECT SAMPLE .000 06:14  
BLANK TEST .000 06:15  
ALCO TO 30.04 in .081 06:16  
BLANK TEST .000 06:17

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

APRIL 21, 2010  
TIME 06:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
SEATERS: 1  
SAMPLE CHAMBER: 42c  
BREATH TUBE: 42c  
BAROMETER: 30.04 in  
FLAM DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z [ \ ] ^ \_ `